

INVITATION FOR QUOTATION

TEQIP-II/2016/CH1G02/Shopping/ 145

10-JUNE-2016

To,

Sub: Invitation for Quotations for supply of Goods

Dear Sir,

1. You are invited to submit your most competitive quotation for the following goods with item wise detailed specifications given at Annexure I,

Sr. No	Brief Description	Quantity	Delivery Period(In days)	Place of Delivery	Installation Requirement (if any)
1	Thermohygrometer	5 UNITS	40	Dr SS Bhatnagar University Institute of Chemical Engg and Tech, Panjab University, Chandigarh	Will be done at the site
2	Digital Techometer	3 UNITS	40	Dr SS Bhatnagar University Institute of Chemical Engg and Tech, Panjab University, Chandigarh	Will be done at the site
3	Stop watch	10 UNITS	40	Dr SS Bhatnagar University Institute of Chemical Engg and Tech, Panjab University, Chandigarh	Will be done at the site
4	Weighing Machine	2 UNITS	40	Dr SS Bhatnagar University Institute of Chemical Engg and Tech, Panjab University, Chandigarh	Will be done at the site

2. Government of India has received a credit from the International Development Association (IDA) towards the cost of the **Technical Education Quality Improvement Programme[TEQIP]-Phase II** Project and intends to apply part of the proceeds of this credit to eligible payments under the contract for which this invitation for quotations is issued.
3. Quotation,
 - 3.1 The contract shall be for the full quantity as described above.
 - 3.2 Corrections, if any, shall be made by crossing out, initialing, dating and re writing.
 - 3.3 All duties and other levies payable by the supplier under the contract shall be included in the unit price.
 - 3.4 Applicable taxes shall be quoted separately for all items.
 - 3.5 The prices quoted by the bidder shall be fixed for the duration of the contract and shall not be subject to adjustment on any account.
 - 3.6 The Prices should be quoted in Indian Rupees only in the prescribed format enclosed at Annexure II.
4. Each bidder shall submit only one quotation.
5. Quotation shall remain valid for a period not less than **55** days after the last date of quotation submission.
6. Evaluation of Quotations,

The Purchaser will evaluate and compare the quotations determined to be substantially responsive i.e. which

 - 6.1 are properly signed ; and
 - 6.2 confirm to the terms and conditions, and specifications.
7. The Quotations would be evaluated for all items together.
8. Award of contract:

The Purchaser will award the contract to the bidder whose quotation has been determined to be substantially responsive and who has offered the lowest evaluated quotation price.

 - 8.1 Notwithstanding the above, the Purchaser reserves the right to accept or reject any quotations and to cancel the bidding process and reject all quotations at any time prior to the award of contract.
 - 8.2 The bidder whose bid is accepted will be notified of the award of contract by the Purchaser prior to expiration of the quotation validity period. The terms of the accepted offer shall be incorporated in the purchase order.
9. Payment shall be made in Indian Rupees as follows:

Delivery and Installation - 90% of total cost

Satisfactory Acceptance - 10% of total cost

10. All supplied items are under warranty of **12** months from the date of successful acceptance of items.
11. You are requested to provide your offer latest by **14:30** hours on **27-JUNE-2016**.
12. Detailed specifications of the items are at Annexure I.
13. Training Clause (if any): The company should provide training of the specified equipment at the time of installation without any additional cost. Also training of concerned person/technician on site.
14. Testing/Installation Clause (if any): Vendor has to install the equipment in the lab space provided at his own cost and test the equipment.
- 15. Information brochures/ Product catalogue is mandatory and accompanied with the quotation clearly indicating the model quoted for.**
- 16. The vendor has to provide the detailed list of institutions/companies (in India) where he has supplied the items along with the list, year of installation and contact person details.**
17. Each bidder will submit bid under two bid system as follow:-

(a) First Envelop will contain:

- i. Technical bid confirming specification, as per Annexure-I
- ii. Information Brochure's/ Product catalogue clearly indicating the model quoted for
- iii. Detailed list of institutions/companies (in India) where you have supplied these items along with the list and year of installation and contract person in details

(b) Second Envelop will Contain:

Financial bid indicating items wise price for items mentioned in the technical bid on the format at Annexure-II.

The Vendor will submit the quotation in above envelopes duly sealed. On first Envelop the vendor will mention Technical bid and on second envelop the vendor will mention price bid. Both these sealed envelope will be put in one closed envelope super scribing clearly the reference of NIQ and the date of opening of bid. The first envelope will be opened first After Evaluation of Technical bid, the second Envelop i.e. price bid will be opened only of those vendors who are meeting the technical specification.

18. Sealed quotation to be submitted/ delivered at the address mentioned below,

**The Chairperson,
TEQIP-II,
Dr. S.S. Bhatnagar University Institute of Chemical Engineering and Technology,
Panjab University,
Chandigarh – 160 014**

18. We look forward to receiving your quotation and thank you for your interest in this project.

(Authorized Signatory)

Name & Designation

Annexure I

Sr. No	Item Name	Specifications																																				
1	Thermohygrometer	<table border="1" data-bbox="613 262 1494 625"> <tr> <td>Measuring rate</td> <td>18s</td> </tr> <tr> <td>Storage temp.</td> <td>-40 to -70⁰ C</td> </tr> <tr> <td>Battery type</td> <td>9 V Block Battery</td> </tr> <tr> <td>Battery life</td> <td>Approx. 1year</td> </tr> <tr> <td>Weight</td> <td>168g</td> </tr> <tr> <td>Dimensions</td> <td>111 X 90 X 40 mm</td> </tr> <tr> <td>Material/ housing</td> <td>ABS</td> </tr> <tr> <td>Display</td> <td>LCD, 2lines</td> </tr> <tr> <td>Warranty</td> <td>2 YEARS</td> </tr> </table> <table border="1" data-bbox="613 688 1494 1003"> <thead> <tr> <th colspan="3">Technical data</th> </tr> <tr> <th>Sensor type</th> <th>NTC</th> <th>Testo humid. Sensor, cap</th> </tr> </thead> <tbody> <tr> <td>Meas. range</td> <td>0 to 50⁰ C -20 to +50⁰ C td</td> <td>+10 to +95 % RH</td> </tr> <tr> <td>Accuracy ±1 digit</td> <td>±0.5⁰ C (at +25⁰ C)</td> <td>±% RH (10 to 95% RH)</td> </tr> <tr> <td>Resolution</td> <td>0.1⁰ C</td> <td>0.1%RH</td> </tr> <tr> <td>Oper.temp.</td> <td colspan="2">0 to 50⁰ C</td> </tr> </tbody> </table>	Measuring rate	18s	Storage temp.	-40 to -70 ⁰ C	Battery type	9 V Block Battery	Battery life	Approx. 1year	Weight	168g	Dimensions	111 X 90 X 40 mm	Material/ housing	ABS	Display	LCD, 2lines	Warranty	2 YEARS	Technical data			Sensor type	NTC	Testo humid. Sensor, cap	Meas. range	0 to 50 ⁰ C -20 to +50 ⁰ C td	+10 to +95 % RH	Accuracy ±1 digit	±0.5 ⁰ C (at +25 ⁰ C)	±% RH (10 to 95% RH)	Resolution	0.1 ⁰ C	0.1%RH	Oper.temp.	0 to 50 ⁰ C	
Measuring rate	18s																																					
Storage temp.	-40 to -70 ⁰ C																																					
Battery type	9 V Block Battery																																					
Battery life	Approx. 1year																																					
Weight	168g																																					
Dimensions	111 X 90 X 40 mm																																					
Material/ housing	ABS																																					
Display	LCD, 2lines																																					
Warranty	2 YEARS																																					
Technical data																																						
Sensor type	NTC	Testo humid. Sensor, cap																																				
Meas. range	0 to 50 ⁰ C -20 to +50 ⁰ C td	+10 to +95 % RH																																				
Accuracy ±1 digit	±0.5 ⁰ C (at +25 ⁰ C)	±% RH (10 to 95% RH)																																				
Resolution	0.1 ⁰ C	0.1%RH																																				
Oper.temp.	0 to 50 ⁰ C																																					
2	Digital Techometer	<p>General technical data Operating temperature 0 to +50 °C Storage temperature -20 to +70 °C Battery type 2 AA batteries Battery life 40 h Display 5-figure LCD display, 1-line Dimensions 175 x 60 x 28 mm Weight 190 g</p> <p>Sensor types Optically with mod. light beam Meas. range +1 to +99999 rpm Accuracy ±1 digit : ±0.02% of mv ±0.02% of mv Resolution : 0.01 rpm (+1 to +99.99 rpm) 0.1 rpm (+100 to +999.9 rpm) 1 rpm (+1000 to +99999 rpm)</p> <p>Mechanical : Meas. Range : +0.1 to +19999 rpm Accuracy ±1 digit : 0.01 rpm (+1 to +99.99 rpm) 0.1 rpm (+100 to +999.9 rpm)</p>																																				

		1 rpm (+1000 to +99999 rpm)																																				
		<table border="1" style="width: 100%; border-collapse: collapse;"> <thead> <tr> <th></th> <th style="text-align: center;">0.1 m</th> <th style="text-align: center;">6"</th> <th style="text-align: center;">12"</th> </tr> </thead> <tbody> <tr> <td style="text-align: center;">m/min</td> <td style="text-align: center;">0.10- 1999</td> <td style="text-align: center;">0.10- 1524</td> <td style="text-align: center;">0.40- 609.6</td> </tr> <tr> <td style="text-align: center;">ft/min</td> <td style="text-align: center;">0.40- 6550</td> <td style="text-align: center;">0.40- 5000</td> <td style="text-align: center;">0.40- 2000</td> </tr> <tr> <td style="text-align: center;">in/min</td> <td style="text-align: center;">4.00- 78700</td> <td style="text-align: center;">4.00- 60000</td> <td style="text-align: center;">4.00- 24000</td> </tr> <tr> <td style="text-align: center;">m/sec</td> <td style="text-align: center;">0.10- 33.30</td> <td style="text-align: center;">0.10- 25.40</td> <td style="text-align: center;">0.10- 10.16</td> </tr> <tr> <td style="text-align: center;">ft/sec</td> <td style="text-align: center;">0.10- 109</td> <td style="text-align: center;">0.10- 83.33</td> <td style="text-align: center;">0.10- 33.33</td> </tr> <tr> <td style="text-align: center;">m</td> <td style="text-align: center;">0.00- 99999</td> <td style="text-align: center;">0.00- 99999</td> <td style="text-align: center;">0.00- 99999</td> </tr> <tr> <td style="text-align: center;">ft</td> <td style="text-align: center;">0.00- 99999</td> <td style="text-align: center;">0.00- 99999</td> <td style="text-align: center;">0.00- 99999</td> </tr> <tr> <td style="text-align: center;">in</td> <td style="text-align: center;">0.00- 99999</td> <td style="text-align: center;">0.00- 99999</td> <td style="text-align: center;">0.00- 99999</td> </tr> </tbody> </table>		0.1 m	6"	12"	m/min	0.10- 1999	0.10- 1524	0.40- 609.6	ft/min	0.40- 6550	0.40- 5000	0.40- 2000	in/min	4.00- 78700	4.00- 60000	4.00- 24000	m/sec	0.10- 33.30	0.10- 25.40	0.10- 10.16	ft/sec	0.10- 109	0.10- 83.33	0.10- 33.33	m	0.00- 99999	0.00- 99999	0.00- 99999	ft	0.00- 99999	0.00- 99999	0.00- 99999	in	0.00- 99999	0.00- 99999	0.00- 99999
	0.1 m	6"	12"																																			
m/min	0.10- 1999	0.10- 1524	0.40- 609.6																																			
ft/min	0.40- 6550	0.40- 5000	0.40- 2000																																			
in/min	4.00- 78700	4.00- 60000	4.00- 24000																																			
m/sec	0.10- 33.30	0.10- 25.40	0.10- 10.16																																			
ft/sec	0.10- 109	0.10- 83.33	0.10- 33.33																																			
m	0.00- 99999	0.00- 99999	0.00- 99999																																			
ft	0.00- 99999	0.00- 99999	0.00- 99999																																			
in	0.00- 99999	0.00- 99999	0.00- 99999																																			
3	STOP WATCH	<p>Requires for all undergraduate experiments in fluid flow, heat transfer and mass transfer to measure time during the experiment</p> <p>Electronic: Reading: 1/100th of second with alarm/timer/calendar display Mechanical: Reading: 1/5th or 1/10th of second with 1/2/3 button</p>																																				
4	Weighing Machine	<p>Platform Scale Capacity : 300kg Readability : 20 / 50 g Pan Size : 600 x 600 / 23.6" x 23.6" Certification: ISO 9001: 2008</p>																																				

Annexure-II

FORMAT FOR QUOTATION SUBMISSION

(In letterhead of the supplier with seal)

Date: _____

To:

Sl. No.	Description of goods (with full Specifications)	Qty.	Unit	Quoted Unit rate in Rs. (Including Ex Factory price, excise duty, packing and forwarding, transportation, insurance, other local costs incidental to delivery and warranty/ guaranty commitments)	Total Price (A)		Sales tax and other taxes payable	
					In figures	In words	In %	In figures (B)
Total Cost								

Gross Total Cost (A+B): Rs. _____

We agree to supply the above goods in accordance with the technical specifications for a total contract price of Rs. _____ (Amount in figures) (Rupees _____ amount in words) within the period specified in the Invitation for Quotations.

We confirm that the normal commercial warranty/ guarantee of _____ months shall apply to the offered items and we also confirm to agree with terms and conditions as mentioned in the Invitation Letter.

We hereby certify that we have taken steps to ensure that no person acting for us or on our behalf will engage in bribery.

Signature of Supplier

Name: _____

Address: _____

Contact No: _____